

JOLIET JUNIOR COLLEGE
REQUEST FOR QUOTATION
DEIONIZED WATER SERVICE

Joliet Junior College seeks quotes for deionized water service as outlined below for a three year period from January 2021 through December 2023. The awarded vendor will be required to honor pricing provided through this bid process for the remainder of the contract. Additional contracts for deionized water services will be renewing throughout the year.

Main Campus, 1215 Houbolt Road, Joliet, Illinois 60431

BLDG ERM – Room #2011

1.2 CF Two-bed DI System – consists of one strong acid cation tank and one strong base anion tank, exchanged every 6 months

1.2 CF Mixed bed DI tank, replaced as needed or one time per year

2ND FLOOR PREP AREA

5UM 10” prefilter, no gasket, replaced every 6 months

GAC 20” 20UM Carbon Filter Cartridge, replaced every 6 months

1.2 CF Tribed DI System – containing one strong acid cation tank, one strong base anion tank, and one mixed bed tank, replaced every 6 months

Romeoville Campus, 1125 135th Street, Romeoville, Illinois 60446

RMB – Room #1045 – Lab DI

.45 CF Mixed Bed Service Exchange Deionization Tank, replaced as needed or every 6 months

10” Carbon Filter Cartridge, replaced every 6 months

10” – 0.2UM Microbial Filter, exchanged every 6 months

RMB – Room #1046 – Lab DI

.45 CF Mixed Bed Service Exchange Deionization Tank, replaced as needed or every 6 months

10” Carbon Filter Cartridge, replaced every 6 months

10” – 0.2UM Microbial Filter, exchanged every 6 months

RMB – Room #1064 – Lab DI

.45 CF Mixed Bed Service Exchange Deionization Tank, replaced as needed or every 6 months

10” Carbon Filter Cartridge, replaced every 6 months

10” – 0.2UM Microbial Filter, exchanged every 6 months

Bidders must also provide a per room cost for the installation of these systems, if applicable.